

Bid Document

Bid Details	
Bid End Date/Time	25-03-2019 17:00:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Delhi
Department Name	Health And Family Welfare Department Delhi
Organisation Name	Dr.b.r.sur Homeopathic Medical College Hospital And Research Centre
Office Name	Principal's Office
Total Quantity	1
Item Category	Defibrillator with CPR Monitoring and TC Pacing
Average Turnover (Last 3 Years)	2 Lakh (s)
OEM Average Turnover (Last 3 Years)	2 Lakh (s)
Experience with Government Required	5 Year (s)
MSE Exemption For Years Of Experience And Turnover	Yes
Startup Exemption For Years Of Experience And Turnover	Yes

Defibrillator With CPR Monitoring And TC Pacing (1)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Product configuration and features	Defibrillator should be light weight, small size with bright coloured display:	Yes	*
	Display screen size (diagonal)	7	More than 7.0 Or higher
	Defibrillator should be Biphasic waveform with 3 waveform display	Yes	*

Defibrillator should display selected and delivered energy, both	Yes	*
Ability to select energy, from paddles as well as from the Defibrillator unit	Yes	*
Manual mode of the unit should provide energy selection at:	1 to 10, 15, 20,30, 50, 70, 85, 100, 150, 200 joules	*
Have ability to measure chest compression rate and depth in real time with both visual and audible feedback	Yes	*
Ability to display CPR index on screen real time:	Yes	Yes, No
Have transcutaneous external pacing with 40 mili second pulse width	Yes	*
Unit should do set test daily with facility to give print out of defibrillator testing report and also have code ready indicator on unit	Yes	*
Ability to filter out CPR artefacts and allowing person to see organised rhythms without interrupting chest compression	Yes	*
Capability to monitor Parameter(s):	ECG	ECG
Capability of internal defibrillation, if and when required?	No	Yes, No
Accessories to be supplied with each unit of defibrillator:	Li-ion battery - 1 no., NIBP paediatric/adult cuff with hose - 1 no. each, Reusable airway adapter to be used with ETCO2 mainstream sensor & cable - 1 no., Multi-function defibrillator/Pacing pads - 24 nos., Reusable CPR feedback sensor/or similar product reused at least	*

		on 90 patients - 2 nos.	
Certification & Warranty	Certification	US FDA approved	*
	Warranty	3	3, 5 Or higher

* Bold specifications are the golden parameters.

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Prabir Manna	110021,DR. B.R. Sur Homoeopathic Medical College Hospital & Research Centre, Nanakpura, Moti Bagh, New Delhi	1	15

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Buyer Specific Additional Terms and Conditions

- 1. Scope of supply (Bid price to include all cost components) :** Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
- 2. Procurement under this bid is reserved for purchase from :** Micro and Small Enterprises from the State of Bid Inviting Authority whose credentials are validated online through Udyog Aadhaar for that product category.
- 3. Purchase preference will be given to MSME and Startup (Local supplier) :** (if L-1 is not from selected category and 4 / 5 Seller (s) from selected category has/have quoted price within L-1 + Micro and Small Enterprises from the State of Bid Inviting Authority whose credentials are validated online through Udyog Aadhaar for that product category (Preference to MSME - 15%) % of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and award contract for percentage of total value as defined/ decided in relevant policy) Micro and Small Enterprises from the State of Bid Inviting Authority whose credentials are validated online through Udyog Aadhaar for that product category (Preference to MSME - 15%)
- 4. Inspection charges shall be borne by the supplier and it shall not be reimbursed by indenter to them at any point in time. Cost of inspection shall be factored in by the supplier in the cost of the product itself while submitting bid. The inspection agency will conduct the inspection and submit the certificate to the supplier as well as the buyers and consignees. The bidder will be solely responsible for pre-inspected material for Quality Assurance including pre-despatch inspection/ testing at manufacturer's site by NABCB SGS India Pvt. Ltd.**

5. Special Terms and Conditions as defined by world bank at [click here](#) will also be applicable

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---