

## CC2 COLUMN PERFORMANCE

Chromatographic condition:

Mobile phase : (70:30) Methanol: Water  
 Flow rate : 1ml/min  
 Column thermostat : 25deg  
 Injection volume : 10µl  
 Detection wavelength : 254nm  
 Test Mix: elution order- Uracil, Methyl benzoate, Toluene and Napthalene

CC2 Column

